Introduction

Name: ZL-2720 QDS-15KV Enameled Wire Voltage Tester



•Implementation standard: IEC60851-5.

 Tel: <u>+86-769-88055358</u>
 Fax: <u>+86-769-88055357</u>

 Website: <u>www.zltester.com</u>
 E-mail: <u>Zhongli@zltester.com</u>

Address: Room 722, Dongyuan Mansion, No.159 QiFeng Road, Dongguang city, P.R.C

- •It is provided with three voltage-rising speeds: 20V/S, 100V/S, and 500V/S
- •It can test five samples at one time
- •LED can display test results, and return to zero position automatically
- •When the sample is cut through, its voltage will keep displaying for from 2 to 10 seconds (adjustable), which will make the testing results clear and distinguishable
- •It is equipped with testing clamps and poises for round stick method and wring method
- •Safety precautions of this tester are in line with relevant provisions of high voltage test equipments, and its door has safe interlock
- •It also can be applied to withstand voltage test.

Models for your choice: QDS-20KV Enameled Wire Voltage Tester;

QDS-15KV-B Enameled Wire Voltage Tester (single electrode)

Cutting-through Current	5±0.5mA
Highest Testing Voltage	≥15000V
Error of Voltage Displayed	<2%
Rated Capacity of Testing Transformer	1000VA
Diameter of Round Stick's Electrode	25±0.5mm
Time of Voltage Withstand	1s~99min99s preset
Work Environment	Temperature: 0~40°C
	Relative humidity: ≤85%
Input Power Supply	AC220V±5% 50Hz
Consumed Power	≤1200W
Dimension	L×W×H 550×600×1700mm
Weight	120Kg

 Tel: +86-769-88055358
 Fax: +86-769-88055357

 Website: www.zltester.com
 E-mail: Zhongli@zltester.com

Address: Room 722, Dongyuan Mansion, No.159 QiFeng Road, Dongguang city, P.R.C



 Tel: +86-769-88055358
 Fax: +86-769-88055357

 Website: www.zltester.com
 E-mail: Zhongli@zltester.com

Address: Room 722, Dongyuan Mansion, No.159 QiFeng Road, Dongguang city, P.R.C